

IMPORTANT NOTICE

Drugs of Abuse II Urine Array

Catalogue Numbers

EV3655 **EV3654** **EV3662**

Please note the crossreactivity section for Buprenorphine has been updated.

The Buprenorphine assay has 21% crossreactivity
with Buprenorphine Glucuronide.

If there are any further queries please contact Biochip Customer Support on
evsupport@radox.com or investigatorsupport@radox.com

Dr. Pauline Armstrong
Quality Assurance Manager

28th May 10