

IMPORTANT NOTICE

Thyroid Total & Free Evidence and Investigator Kits

EV3501, EV3502, EV3545, EV3546, EV3516 & EV3517

For the Evidence Thyroid Arrays the tolerance ranges assigned for the quality control DTRs have been amended for improved performance.

Please ensure that the update disc is run prior to analysing samples.

If there are any further queries please contact Biochip Customer Support at evidencesupport@randox.com or investigatorsupport@randox.com

Dr. Pauline Armstrong
Quality Assurance Manager

7th July 10

PCA 3285