

# IMPORTANT NOTICE

Cytokine I Array

Catalogue Number CY5006

Following an internal review of the cytokine array it was noted that control ranges for the following lots required re-assignment.

Lot 4219EV (CY5006 batch 3085)  
Lots 4564-4566EV (CY5006 batch 3318)

A replacement control update disc will be provided through Biochip Customer Support.

If there are any further queries please contact Biochip Customer Support at [investigatorsupport@randox.com](mailto:investigatorsupport@randox.com) or [evidencesupport@randox.com](mailto:evidencesupport@randox.com)

Dr. Pauline Armstrong  
Quality Assurance Manager

5<sup>th</sup> July 2010